

# UNIVERSITY OF BIRMINGHAM

## Research at Birmingham

### Statistical Description of Inhomogeneous Samples by Scanning Microwave Microscopy

Monti, Tamara; Udoudo, Ofonime; Sperin, Kevin; Dodds, Chris; Jackson, Timothy; Kingman, Sam

DOI:

[10.1109/TMTT.2016.2642940](https://doi.org/10.1109/TMTT.2016.2642940)

License:

Other (please specify with Rights Statement)

*Document Version*

Peer reviewed version

*Citation for published version (Harvard):*

Monti, T, Udoudo, O, Sperin, K, Dodds, C, Jackson, T & Kingman, S 2017, 'Statistical Description of Inhomogeneous Samples by Scanning Microwave Microscopy', IEEE Transactions on Microwave Theory and Techniques, vol. 65, no. 6, pp. 2162-2170. <https://doi.org/10.1109/TMTT.2016.2642940>

[Link to publication on Research at Birmingham portal](#)

#### **Publisher Rights Statement:**

(c) 2017 IEEE. Personal use of this material is permitted. Permission from IEEE must be obtained for all other users, including reprinting/republishing this material for advertising or promotional purposes, creating new collective works for resale or redistribution to servers or lists, or reuse of any copyrighted components of this work in other works

#### **General rights**

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

#### **Take down policy**

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact [UBIRA@lists.bham.ac.uk](mailto:UBIRA@lists.bham.ac.uk) providing details and we will remove access to the work immediately and investigate.

# Statistical Description of Inhomogeneous Samples by Scanning Microwave Microscopy

Tamara Monti, Ofonime B. Udoudo, Kevin A. Sperin, Chris Dodds, Sam W. Kingman, and Timothy J. Jackson

**Abstract**—A quantitative analysis of the dielectric properties of a multiphase sample using a scanning microwave microscope (SMM) is proposed. The method is demonstrated using inhomogeneous samples composed of a resin containing micrometric inclusions of a known ceramic material. The SMM suitable for this task employs relatively large tips (tens of micrometers in diameter). Additionally, in order to make the instrument more suitable for high-throughput analysis, an original design for rapid tip changes is implemented. Single-point measurements of dielectric constant at random locations on the sample were performed, leading to histograms of dielectric constant values. These are related to the dielectric constants of the two phases using Maxwell–Garnett effective medium theory, taking into account the volume-of-interaction in the sample beneath the tip.

**Index Terms**—Dielectric constant, dielectric materials, Maxwell–Garnett approximation, microwave measurements, near-field measurements, nonhomogeneous media, scanning microwave microscope (SMM), statistical distributions.

## I. INTRODUCTION

THE analysis of the dielectric properties of multiphase materials is important in many scientific fields [1]. The near-field scanning microwave microscope (SMM) has been extensively exploited for this purpose [1], even though the problem of modeling the tip–sample interaction is still an open research topic [2], [3].

Many implementations of the SMM are based on atomic-force microscope platforms, using noncontact interaction with very fine (nm scale) tips [4]. Such approaches are not suitable for rough surfaces or large samples where the features of interest may be many tens of micrometers in extent. In this case, coarser movements and larger tips are appropriate. If considering, for example, circuits and integrated devices for power electronics, the single elements and connections are usually several hundreds of micrometers in size [5]. These features are still too small to be efficiently tested by a conventional miniaturized dielectric probe [6]. A similar situation is found in other applications, ranging from the high-power microwave processing of materials [7], where the dielectric features of micrometric inclusions in the treated materials strongly influence the treatment efficiency itself [8], to the diagnosis

of 3-D-printed circuit elements [9], especially when assembled in a complex board. Engineering applications like these usually involve natural or man-made multiphase materials at micrometric scale, with uncontrolled roughness and partially known composition. So far, the SMM has hardly been used to study such materials, even though potential benefits, such as fault detections and process quality assessments, can come from dielectric analysis.

Additionally, in realistic contexts, such as *in situ* high-throughput diagnostic processes, a full scan of the object under analysis is not required, and would be extremely time-consuming. A statistical description of the dielectric properties is much more suitable, provided some *a priori* information about the materials under analysis. This can be applied in the case of a production chain where the location of the assembled components is known and defects can be identified by the variations from the known topology.

The ability to rapidly change tips is essential to such application of SMM. Furthermore, in such cases, a semiempirical approach for calibration is pragmatic to use. Such a calibration provides validation of the instrument, it does not require detailed knowledge of the geometry of the tip, and it yields simply derived estimates of the error bars.

In this paper, a bespoke SMM with the features described above (tips of tens of micrometers size, and agile system for changing tips) is applied to the dielectric characterization of a set of exemplar two-phase samples. A statistical strategy is adopted for a comprehensive description of the samples' dielectric properties. Rather than collecting sample images, the SMM dielectric measurements are taken as single points at locations chosen randomly. The dielectric permittivity results are related to the distribution of one phase with respect to the other by considering the Maxwell–Garnett approximation for the effective medium included in the SMM volume of interaction [10]. Such volume extends in all the three planes defined by  $x$ ,  $y$ , and  $z$ . The SMM probe is insensitive to the material properties outside this volume. Due to the random nature of the SMM measurements, the composition of the portion of the sample under analysis is not known. In order to confirm the validity of the effective medium approximation, the samples are subsequently analyzed by the scanning electron microscope – extended backscattering electron (SEM-XBSE) technique interpreted by the Mineral Liberation Analysis (MLA) software. The MLA provides an automated scan of the sample surface with pattern matching of the X-rays collected in a database. The software is used to build an image from which particle parameters, including dimensions, are calculated. This analysis is necessary in this paper in order to precisely determine the area of the inclusions (first phase)

Manuscript received September 1, 2016; revised November 18, 2016 and December 16, 2016; accepted December 18, 2016.

T. Monti, O. B. Udoudo, C. Dodds, and S. W. Kingman are with the Faculty of Engineering, University of Nottingham, Nottingham NG7 2RD, U.K. (e-mail: tamara.monti@nottingham.ac.uk; ofonime.udoudo@nottingham.ac.uk; chris.dodds@nottingham.ac.uk; sam.kingman@nottingham.ac.uk).

K. A. Sperin and T. J. Jackson are with the School of Engineering, University of Birmingham, Birmingham B15 2TT, U.K. (e-mail: kxs023@student.bham.ac.uk; t.j.jackson@bham.ac.uk).

Color versions of one or more of the figures in this paper are available online at <http://ieeexplore.ieee.org>.

Digital Object Identifier 10.1109/TMTT.2016.2642940

and the ratio of these with respect to the host medium (second phase).

The theoretical methods used to evaluate the performance of the SMM and analyze the data are presented in Section II. Section III details the experimental procedures. The results are presented and discussed in Sections IV and V, respectively. The novel features of this paper are a determination of the lower bounds of dielectric loss that can be resolved by such an SMM (Section IV-A), the statistical treatment of two-phase samples with independent characterization of physical properties (Section IV-C), the consideration of the volume of interaction beneath the SMM tip, and, from the design point of view, the SMM head that enables a rapid interchange of probe tips (Section III).

The final aim of this paper is to widen the application of the SMM to new areas where it is important to have high-throughput dielectric information of relatively large samples and with micrometric inclusions. The practical novelties of the proposed design and the statistical approach (Section III), instead of full scans of the samples, make the proposed instrument suitable for the latter purpose. The semiempirical model used for deriving the dielectric constant from the microwave measurements, described in Section II, makes the technique accessible outside the microwave community as well.

## II. THEORY

In most cavity resonator SMMs, including the one used in this paper, a sharp tip protrudes through the aperture from the center conductor. This is commonly modeled as a capacitance  $C_t$  connected in parallel with the resonator capacitance  $C$ . A sample beneath the tip is modeled by an additional impedance [11]–[13] related to the sample's complex relative permittivity  $\varepsilon_1 - j\varepsilon_2$ . The sample may be considered “low loss” if its loss tangent  $\tan \delta = \varepsilon_2/\varepsilon_1 \gg Q_L$ , where  $Q_L$  is the loaded quality factor of the resonator, because in this case the resonator will be insensitive to dissipation in the sample. Such a sample presents only an additional capacitance  $C_s$  in series with  $C_t$ . The resonance frequency of the probe will fall from  $f_0$  without the sample to a lower value  $f$  with the sample. This is expressed as a normalized frequency shift of

$$\frac{\Delta f}{f_0} = -\frac{1}{2} \frac{C_t C_s}{C(C_t + C_s)} \quad (1)$$

when the shift is small. Equation (1) needs a model for  $C_s$  as a function of sample permittivity to be really useful. Alternative estimates of the frequency shift [14], [15] have been obtained by calculation of the electric field around spherical [14] and axisymmetric [15] tips using an image charge analysis. The frequency shift was calculated using perturbation theory in the spherical case to be

$$\frac{\Delta f}{f} \approx \frac{\Delta f}{f_0} = A \left( \frac{\ln(1-b)}{b} + 1 \right) \quad (2)$$

with  $b = (\varepsilon_1 - 1)/(\varepsilon_1 + 1)$ . An alternative empirical expression has also been proposed [16]

$$\frac{\Delta f}{f} = -\Gamma \frac{\varepsilon_1 - 1}{\varepsilon_1 + K}. \quad (3)$$

In this model,  $\Gamma$  and  $K$  are constants principally related to the distribution of electric fields around the tip and the size of the aperture, respectively. All the three models are based on the assumption that the current distribution within the resonator is not changed by the sample. Both (2) and (3) can be arranged in a linear form, yielding gradient and intercept terms that become the calibration constants. The  $x$ -ordinate data should be taken with respect to the mean value in order to remove the correlation and covariance between the gradient and the intercepts [17].

Once the calibration data have been obtained, yielding gradient  $m$  and intercept  $c$ , frequency shift data from an unknown sample can be converted to relative permittivity via an inversion. For the Inoue model (3), the appropriate equation is

$$\varepsilon_1 = \frac{1 + (m\bar{\varepsilon}_1 - c)(\Delta f/f)}{1 + m(\Delta f/f)} \quad (4)$$

where  $\bar{\varepsilon}_1$  is the mean value of  $\varepsilon_1$  in the calibration set. The errors in the calibration data can be estimated by assuming the model to be a good description of the data and computing the standard deviation  $\sigma$  required to make

$$\chi^2 = \sum_i \frac{(y_i - (\Delta f/f)_i)^2}{\sigma^2} = 1. \quad (5)$$

The errors in the gradient and intercept parameters can be calculated using the value of  $\sigma$  deduced from

$$\sigma_m = \sqrt{\frac{\sigma^2}{N(\bar{\varepsilon}_1^2 - \bar{\varepsilon}_1^2)}} \quad (6)$$

and

$$\sigma_c = \sqrt{\frac{\sigma^2 \bar{\varepsilon}_1^2}{N(\bar{\varepsilon}_1^2 - \bar{\varepsilon}_1^2)}} \quad (7)$$

where  $N$  is the number of samples in the set. The error terms in (6) and (7) may be used to calculate the error bars for the extracted relative permittivity  $\varepsilon_1$ . Since the gradient  $m$  and intercept  $c$  are independent, (8) can be used to derive the experimental error in the determination of  $\varepsilon_1$  as

$$\sigma_{\varepsilon_1} = \sqrt{\left(\frac{d\varepsilon_1}{dm}\right)^2 \sigma_m^2 + \left(\frac{d\varepsilon_1}{dc}\right)^2 \sigma_c^2}. \quad (8)$$

The tip-sample interaction in the two-phase media considered in this paper can be modeled using the Maxwell–Garnett approximation. The relative dielectric permittivity of the effective medium interacting with the tip is [18], [19]

$$\varepsilon_1^{eff} = \varepsilon_1^h + \frac{3V\varepsilon_1^h}{\frac{\varepsilon_1^i + 2\varepsilon_1^h}{\varepsilon_1^i - \varepsilon_1^h} - V} \quad (9)$$

where  $\varepsilon_1^{eff}$ ,  $\varepsilon_1^h$ , and  $\varepsilon_1^i$  are the relative dielectric permittivity of the effective medium, the host medium, and of the inclusions, respectively.  $V$  is the volume fraction filled by the inclusions within the total volume of interaction. The Maxwell–Garnett approximation requires that the working




Fig. 1. SMM. Samples are placed beneath the resonator on a spring-loaded cantilever platform. Deflection of the platform as the resonator is lowered into contact with the sample is recorded with a distance sensor located beneath the platform. The sample may be scanned by means of the motorized stages shown. Sample movements are performed out of contact with the resonator. Left: schematic of the quarter-wavelength coaxial resonator.

regime is “quasi-static,” which is reasonable here because of the near-field interaction due to the proximity of the probe and sample [19]. It is assumed that the inclusions are physically separated. The limited volume of interaction with respect to the size of the inclusions ensures that the effect of the particle shape is not relevant [18].

In order to model the random variability of the interacting volume between tip and sample for each measurement,  $V$  was considered as a statistical variable with Gaussian distribution  $f(x)$  with a certain mean value  $\mu$  and standard deviation  $\sigma$

$$f(x | \mu, \sigma^2) = \frac{1}{\sigma \sqrt{2\pi}} e^{-\frac{(x-\mu)^2}{2\sigma^2}}. \quad (10)$$

The  $\mu$  and  $\sigma$  parameters depend on sample composition and particles distribution. This assumption is based on the sample preparation procedure, described in the following section.

### III. EXPERIMENTAL METHODS

The full SMM setup is shown in Fig. 1. A schematic of the SMM resonator is presented in the left of the figure. Such a configuration is inspired by the seminal work in [20] and employed in [1], [14], and [16]. The SMM used here is based on that reported elsewhere [21], adapted for high throughput.

The top of the center conductor is threaded. The center conductor is fixed to the top-plate from inside the body of the resonator by screwing the threaded end through a tapped hole in the center of the top-plate. The center conductor is preloaded with a tip before insertion. This system allows for the rapid change of the tip in the case of damage during the scan. Fig. 2 shows the bottom of the SMM resonator. The bottom plate comprises of two semicircular 3-D-printed plastic disks, covered with copper tape. The two halves can be closed from the sides rather than from above the tip, thus avoiding possible damage.

Scattering parameters were recorded using a frequency-swept signal from a vector network analyzer (Copper Mountain Technologies TR5048). The VNA was typically set to an IF bandwidth of 300 Hz and to record 201 data points over six resonance bandwidths, giving a frequency resolution




Fig. 2. Base of the SMM resonator. One base half-plate is shown attached to the body of the resonator with the tip protruding through the semicircular opening in the centerline of the base. The zoomed-in-view shows the tip more clearly. The bright spot is the solder joint between the tip and the center conductor of the resonator.

of the order of 50 kHz. This is considerably larger than the frequency setting resolution of the instrument, which is 10 Hz. Sweeps were averaged with an averaging factor of ten. The resonant frequency and quality factor were extracted by fitting to the data sets as described elsewhere [22], [23].

To maintain the samples in contact with the tip, a spring-loaded cantilever platform was used as a sample stage. Below the plate, an infrared distance sensor [Philtec Inc, D63-L] was used to measure the deflection and vibration of the plate, as in [24]. The resonator was lowered toward the sample until the vibrations of the plate just stopped. A maximum contact force of 20  $\mu\text{N}$  was deduced from the deflection of the plate.

Two tips were used for measurement of the samples, a 180  $\mu\text{m}$  diameter spherical tungsten tip on a 500  $\mu\text{m}$  diameter shaft [25] and a sharp gold-plated tungsten tip of nominal final diameter 20  $\mu\text{m}$  [PSPTG20100 from Tech-Specialties Inc., Sandia, USA] (Fig. 3). The larger tip was found to be robust, so calibration measurements were taken during the start and end of measurements with that tip and the final mean calibration was applied to the data. The differences in the two calibrations compared with the mean were of the order of 1% in the gradient term and 0.5% in the intercept term. Variations between data sets for the 20  $\mu\text{m}$  tip were observed to be larger, so this tip was calibrated before measurements of each sample. Analysis of the six individual calibrations compared with their mean showed a maximum-to-minimum variation of 25% in the gradient term and 14% in the intercept term, with no systematic trend between data sets.

Calibrations of the resonator were performed with a set of standard samples detailed in [21], [26], and [27]. All were characterized in a split postresonator [21], [26], except the quartz and  $\text{LaAlO}_3$  for which standard values were assumed [27]. The temperature variations in the laboratory during experimental runs were monitored to be of the order of  $\pm 1$   $^\circ\text{C}$ . The coefficient of thermal expansion of copper is 17  $\text{ppm}^{-1}$  [28].

The dielectric materials were chosen in order to present a wide range of dielectric contrast between inclusions and




Fig. 3. Left: SEM pictures of the tips before mounting on the resonator. Scale bars represent 50  $\mu\text{m}$ . Right: photograph of the sample.

hosting medium. The materials chosen for the three samples were as follows.

- 1) BZT (barium zirconate titanate) with  $\epsilon_1 \approx 29.5$  and Q factor of 40-50000 at 1.8 GHz;
- 2) CTNA (calcium titanate neodymium aluminate/ $\text{CaTiO}_3\text{-NdAlO}_3$ ) with  $\epsilon_1 \approx 45$  and Q factor of 25000 at 2 GHz;
- 3) BNT (barium neodymium titanate/ $\text{BaNd}_2\text{Ti}_4\text{O}_{12}$  with  $\epsilon_1 \approx 80$  and Q factor of 5000 at 2 GHz.

These dielectrics are commonly employed in dielectric resonators and they were manufactured and characterized by Filtronic Comtek (<http://www.filtronic.com/>).

Each of the three samples was made by crushing a block of dielectric material and sieving the particle fragments to control their mean size ( $\sim 250 \mu\text{m}$ ) [class size -300+212]. The particles were embedded in epoxy resin in 25 mm cylindrical molds and left to cure overnight. The cured sample was then polished down to  $\sim 1 \mu\text{m}$  roughness (see Fig. 3). The dielectric permittivity of the epoxy was characterized by SMM and also by an open-ended coaxial probe technique [29].

Each sample was examined with the SMM. A first set of measurements with the 180  $\mu\text{m}$  tip was performed, and the permittivity data extracted from the resonant frequency according to the procedure described previously. Between measurements, the resonator was raised and the sample moved to a new position selected by a random number generator. This method was chosen to avoid any bias during the test. More than 50 SMM measurements were taken for each sample. The procedure was then repeated for the 20  $\mu\text{m}$  tip. A distribution of dielectric constant values was obtained from each set of measurements.

Since the sample is composed of two solid phases (epoxy and dielectric), the volume fraction filled by each one was unknown *a priori*. Therefore, in order to further characterize the samples under study, a surface analysis was made using the XBE technique in a FEI Quanta 600i SEM. Quantitative data




Fig. 4.  $|S_{11}(f)|$  data for the closed resonator, resonator with an aperture in the base plate, and resonator with the 20- $\mu\text{m}$  tip. The resonant frequencies are 1.950221, 1.950232, and 1.679313 GHz, respectively. Note the use of a linear vertical scale.

TABLE I  
EQUIVALENT CIRCUIT PARAMETERS

Resonator	$n_1^2$	$R$ ( $\Omega$ )	$L$ (nH)	$C$ (fF)
Closed	0.67	1.34	251	26.5
With aperture	0.67	1.35	250	26.6
With tip	0.67	1.36	146	61.4

Equivalent circuit parameters determined from one port characterization of the resonator. The tip used was the 20  $\mu\text{m}$  tip. The unloaded quality factors are 2294, 2271 and 1131 for the closed resonator, resonator with aperture and resonator with tip, respectively.

on the number of dielectric particles embedded in resin host, the area of each particle, and the surface occupied by them with respect to the total sample surface was obtained through use of the MLA software. The volume of the sample probed by the backscattered electrons is of the order of a few tenths of a micrometer in diameter [30], so the MLA data are surface specific compared with the SMM data where the interaction volume is expected to be determined largely by the diameter of the tip.

#### IV. RESULTS

##### A. One-Port Calibration of the Resonator

Fig. 4 shows the magnitude of  $S_{11}(f)$  for the closed resonator, the resonator with the aperture in the base plate, and the resonator with the 20  $\mu\text{m}$  tip protruding through the aperture. Values of  $S_{11}(f_0)$ ,  $f_0$ , and  $Q_L$  were determined from fitting the theoretical model described in [22] to the data. Table I shows the equivalent circuit parameters determined from the data. Fig. 4 and Table I show that the reactance of the resonator is strongly perturbed by the addition of the tip but that the aperture required for the tip itself presents only a very small perturbation. This explains the very high quality factor in the resonator used in [16], which had no tip. The tip itself presents little dissipation. However, the quality factor is clearly too low compared to the quality factor of both the sample calibration set and the dielectric samples under test to be able to resolve their dielectric losses. For this reason, dielectric loss will not be considered further in this paper.






Fig. 5. Final mean calibration data in the case of 180- $\mu\text{m}$  tip (top) and 20- $\mu\text{m}$  tip (bottom) for the Gao-Xiang image charge model [14] (2).

TABLE II  
FITTING PARAMETERS—GAO-XIANG MODEL [14]

Tip	$m$	$\sigma_m$	$c$	$\sigma_c$
180 $\mu\text{m}$	0.009015	0.000703	-0.012840	0.000401
20 $\mu\text{m}$	0.004915	0.000438	-0.007355	0.000250

Typical fitting parameters from calibration data for the Gao-Xiang image charge model (2).

### B. Two-Port Calibration of the SMM

In what follows, the data shown are from magnitude and phase measurements of  $S_{21}$  taken after a full two-port calibration. The coupling was much weaker than for the one-port measurements in order to make  $S_{21} \leq -20$  dB. The reference planes were at the connection to the resonator coupling ports.

Fig. 5 shows an example of normalized frequency shift obtained from the calibration sample set for the 180  $\mu\text{m}$  tip and for the 20  $\mu\text{m}$  tip by applying the image charge model (2) to the data. The corresponding fitting parameters are given in Table II. The same data are shown in the case of Inoue's model (3) in Fig. 6 and Table III.

The fractional errors for the two models are very similar in magnitude: the data for the image charge model are consistent with an analysis reported previously on the measurement of thin films [21]. However, in most cases the Inoue model (3) gave slightly lower errors than the Gao-Xiang model [14] (2). The fractional standard errors shown in Table III are approximately 7% in the gradient and 3% in the intercept for both




Fig. 6. Final mean calibration data in the case of 180- $\mu\text{m}$  tip (top) and 20- $\mu\text{m}$  tip (bottom) obtained by applying the Inoue model (3).

TABLE III  
FITTING PARAMETERS—INOUE MODEL

Tip	$m$	$\sigma_m$	$c$	$\sigma_c$
180 $\mu\text{m}$	41.97	2.817	931.4	27.39
20 $\mu\text{m}$	76.45	5.526	1628	53.74

Typical fitting parameters from calibration data for the Inoue model (3).

the tips. The Inoue model is the more convenient one of the two for the analysis that follows because it provides a linear equation for extraction of an unknown permittivity from a measured frequency shift and hence leads to a simpler estimation of the uncertainty in the extracted permittivity. For normalized frequency shifts of  $-0.015$  (180  $\mu\text{m}$  tip) and  $-0.008$  (20  $\mu\text{m}$  tip) in the calibration data, the uncertainty in the extracted permittivity  $\sigma_{\epsilon_1}$  is approximately 1, or 10% of the true relative permittivity.

In order to assess the success of the calibration, an epoxy sample was tested by SMM and by an open-ended coaxial probe technique [29]. The dielectric permittivity of the epoxy as characterized by SMM was  $\epsilon_1^h = 3.2$  with a 6% standard deviation over five single-point SMM measurements on different spatial locations. The error is likely to be related to intrinsic variations of the epoxy sample. The value obtained by the open-ended coaxial probe technique was  $\epsilon_1^h = 3.1$ .

Temperature drift in the laboratory during measurements could have caused drift in the measured resonant frequencies of 17.6 ppm, or about 30 kHz. For the MgO sample of permittivity  $\approx 9.8$ , the frequency shifts were typically 12250 ppm for the 180  $\mu\text{m}$  tip and 7370 ppm for the 20  $\mu\text{m}$  tip, respectively.

TABLE IV  
RESULTS OF THE MLA ANALYSIS

Sample	Total area ( $\mu\text{m}$ )	Particle #	Particle area %
BZT	204068452.41	6067	47%
CTNA	191443025.34	5605	44%
BNT	200722131.00	5679	47%




Fig. 7. XBSE images of the samples: darker areas correspond to epoxy while the brighter areas correspond to ceramic particles. (a) BZT inclusions sample. (b) CTNA inclusions sample. (c) BNT sample. The scale bars represent 2 mm. The average inclusion size is around 200  $\mu\text{m}$ .

TABLE V  
GAUSSIAN DISTRIBUTION PARAMETERS

Tip	$\mu$ (BZT)	$\sigma$ (BZT)	$\mu$ (CTNA)	$\sigma$ (CTNA)	$\mu$ (BNT)	$\sigma$ (BNT)
180 $\mu\text{m}$	0.55	0.12	0.52	0.12	0.55	0.12
20 $\mu\text{m}$	0.65	0.06	0.8	0.06	0.8	0.06

Parameters of the Gaussian distribution used for modelling the volume filling fraction.

The error in resonant frequency introduced by temperature drift for MgO was less than 0.25%.

### C. Statistical Description

The XBSE images of the three composite samples are reported in Fig. 7. The darker parts of the sample (lower atomic number) correspond to the epoxy areas while the brighter ones (higher atomic number) represent the ceramic particles. The MLA results for the three samples under test are summarized in Table IV. The “particle area %” refers to the fraction of surface occupied by the mineral phase with respect to the total sample surface. The statistics of the particle areas are reported in Fig. 8.

The results of the fitting of the dielectric data obtained from the SMM measurements with the 180 and 20  $\mu\text{m}$  tip with those calculated via the numerical procedure described in Section III are reported in Table V.

In the 180  $\mu\text{m}$  case, the three different samples are statistically described by volume filling  $V$  with a monovariate Gaussian distribution as the permittivity data are concentrated around one value. The slightly asymmetric shape of the measured permittivity distribution is followed by the values generated numerically, which confirms that the Maxwell–Garnett approximation is applicable for this kind of tip–sample interaction. The results are plotted in Fig. 9.

The results of the fitting for the 20  $\mu\text{m}$  tip measurements are plotted in Fig. 10. In this case, the permittivity distribution




Fig. 8. Histogram of the distribution of the particle areas from MLA analysis.




Fig. 9. Fittings of the experimental data obtained with 180- $\mu\text{m}$  tip (yellow/dashed columns) with the statistical description based on Maxwell–Garnett effective medium approximation (blue/solid columns).

is concentrated around two peak values. It is clear that the statistical description of the interaction volume (in terms of filling fraction) is more complicated than in the former case, suggesting a bivariate distribution. It is interesting to see that, while the second peak in the distribution shifts toward higher values when the inclusion permittivity increases, the first peak remain fixed around  $\epsilon_1^{eff} \sim 5$  for all the samples. The higher peaks were fit similar to the 180  $\mu\text{m}$  case (Table V).




Fig. 10. Fittings of the experimental data obtained with 20- $\mu\text{m}$  tip (yellow/dashed columns) with the statistical description based on Maxwell–Garnett effective medium approximation (blue/solid columns).

TABLE VI  
COMPARISON BETWEEN THE SIMULATED ADMITTANCE VALUES

Tip	Min- Max (BZT)	Epoxy (BZT)	Min- Max (CTNA)	Epoxy (CTNA)	Min- Max (BNT)	Epoxy (BNT)
20 $\mu\text{m}$	0.047	0.060	0.053	0.066	0.060	0.071
180 $\mu\text{m}$	0.020	0.071	0.023	0.078	0.027	0.085

Comparison between the simulated admittance values for the tip interacting with the three different samples under analysis.

The fitting was performed with a total number of occurrences equal to the actual number of measurements. For this reason, in the latter case (20  $\mu\text{m}$ ) the fittings of the higher peaks are just qualitative. The fittings follow the distribution profile of the measured data although the numbers of values within each bin of the histograms do not match.

#### D. COMSOL Model

A numerical model of the tip–sample interaction was developed using the ACDC module of COMSOL Multiphysics (version 4.4). The volume of interaction beneath the tip is expected to be different for the two tips. In particular, the size and positions of the inclusions were chosen in order to simulate the average MLA results, reported in Table IV. An average size of 200  $\mu\text{m}$  and rectangular section were considered. An image of the electric potential distribution (V) is shown in Fig. 11. For the simulations a reference voltage  $V_{00} = 1$  was applied to the tip while the surrounding shield representing the bottom plate of the resonator was grounded. The ratio between the length of the tip and the diameter of the resonator was chosen to match the real resonator, thus reproducing the combined effects of the electric field leaking through the aperture and the electric field produced by charge on the tip.

The values of the admittance  $Y$  were simulated at the reference port (the tip in this case). Accordingly, the difference




Fig. 11. COMSOL ACDC model of the 180- $\mu\text{m}$  tip–sample interaction: profile of the electric potential distribution for the tip directly above a dielectric particle.

in admittance values was compared between the same two positions for all the simulated samples. These results are given in Table VI under the “Min–max” column headings for each sample and are representative of the different tip–sample interactions for the three dielectric contrasts under analysis.

#### V. DISCUSSION

The simple COMSOL model does not allow a direct derivation of the permittivity values but it is reasonable to assume that the admittances are related to the effective permittivity beneath the tip. The results of the simulations are in accordance with the statistical description of the tip–sample interaction of the previous section and show the physical meaning of such statistical description.

In particular, the differences between the 180 and 20  $\mu\text{m}$  interacting tips are evident: in both the cases the value obtained when the tip is placed between the inclusions (tip shift  $x = 160 \mu\text{m}$  hereby referred to as “Min”) is different from the epoxy alone, even though there is epoxy directly under the tip. This is because the tip–sample interaction is volumetric and inclusions are partially present within the interaction volume. The reason for the absence of permittivity values close to  $\epsilon_1^{eff} = 3.2$  (epoxy permittivity) in the measured data is clear. Even if the tip lands on an epoxy surface, the volume of interaction contains some fraction of the surrounding dielectric inclusions.

Additionally, these simulation results help in understanding why the 20  $\mu\text{m}$  tip showed a large number of values in the bin  $\epsilon_1^{eff} \sim 5$ , constant for the three samples under investigation (Fig. 10). The values in the columns labeled “Epoxy” in Table VI are smaller for the 20  $\mu\text{m}$  tip than for the 180  $\mu\text{m}$ . The smaller tip has a smaller volume of interaction with the sample compared with the larger tip. When the smaller tip lands between inclusions, a greater volume fraction will be filled by epoxy and the volume fraction filled by the surrounding inclusions is less than for the larger tip. The effective permittivity is dominated by the epoxy and is therefore lower. Thus, the first peak of permittivity around 5 can be attributed to the measurements done when the tip






Fig. 12. Admittance value profile obtained by COMSOL ACDC model of the 180- $\mu\text{m}$  (top) and 4- $\mu\text{m}$  (bottom) tip-sample interaction.

“landed” on areas filled by epoxy, and is almost constant for the three dielectric contrasts considered in the experiment. The tails on the right-hand side of the first peaks in Fig. 10 are due to intermediate situations when the tip “landed” on epoxy but close to an inclusion. In order to highlight this effect, a further simulation with a 4  $\mu\text{m}$  diameter tip was performed. The raw output of the 180 and 4  $\mu\text{m}$  tip simulations is shown in Fig. 12 for comparison.

There is no sign of a split distribution of values of relative permittivity in the case of the 180  $\mu\text{m}$  tip (Fig. 9), because the volume of interaction is large enough to always be intercepted by inclusions. The fitting parameters reported in Table V statistically represent the distribution of inclusions. The central value  $\mu$  of the CTNA sample is slightly lower than the others and this is in accordance with the MLA analysis results.

A further interesting observation arises from the second peaks of permittivity in the data from the 20  $\mu\text{m}$  tip. The fitting parameters in this case are influenced by the dielectric contrast, defined as the difference between the relative permittivity of the ceramic and the epoxy. In the lowest contrast sample, the BZT  $\mu = 0.65$ , whereas for the CTNA and BNT  $\mu = 0.8$ .

This effect in the SMM data and COMSOL simulations is consistent with the previous work on dielectric mixtures [31], where it was shown that a dielectric contrast of 10 is an

empirical threshold between different models of the interaction between electromagnetic field and sample. There appears to be a need for a filling factor that takes into account the dielectric contrast as well as the geometrical distribution of inclusions within the volume of interaction beneath the tip of the SMM. Although the samples are geometrically similar, as was shown by the MLA data, the fitting parameters for the second peak of the 20  $\mu\text{m}$  data for the higher permittivity samples differed from that for the lowest permittivity sample. For the BZT data, where the dielectric contrast is below 10,  $\mu = 0.65$  is influenced by the geometry of the sample alone and it is the physical filling factor. In the CTNA and BNT data, where the dielectric contrast is above 10, the filling factor  $\mu = 0.8$  is higher, and contains contributions from both the actual filled volume and the dielectric contrast that further segregates the field into the inclusions.

## VI. CONCLUSION

In this paper, a statistical approach is presented for the high-throughput dielectric characterization of multiphase materials samples. Scanning electron microscopy was used to characterize the topology and composition of multiphase dielectric samples. The experimental results match the effective medium description through the Maxwell–Garnett approximation for the portions of the sample under analysis for each measurement. In particular, by considering the fraction of volume of interaction filled by the inclusion ( $V$ ) as a statistical variable, measurements performed with 180  $\mu\text{m}$  tip are described by a monovariate Gaussian distribution of  $V$ . The results of the measurements performed by 20  $\mu\text{m}$  tip follows a bivariate distribution instead. In light of this statistical interpretation of the experimental results, the effect of the sample inhomogeneity with respect to the volume of interaction is explained.

The Inoue model of the tip-to-sample interaction was found more suitable for extraction of the permittivity than the image charge model of Gao and Xiang [14]. The error bars for the calibration derived from the two models were similar, but the error bars for measurement of unknown permittivity are more straightforward to derive with the Inoue model. A complete uncertainty analysis considering the calibration set and practical factors such as sample positioning, tip-to-sample contact, and the geometry of the resonator around the tip as tips are interchanged is required in the future work.

The SMM was designed for rapid interchange of probe tips. The lower bounds of dielectric loss that can be resolved by such an SMM are evaluated. The combination of a versatile instrument with statistical interpretation of high-throughput measurements allows exhaustive characterization of a relatively large sample in a short time. This new concept for SMM characterization can be potentially applied to applications areas, such as production chains or material handling systems, where multiphase materials are employed and topology and composition information are partially known *a priori*.

## ACKNOWLEDGMENT

The authors would like to thank Y. Li for preparatory work on the SMM during his M.Sc. project, M. Wager for fabrication of the SMM base-plates, Prof. C. C. Constantinou for photography, and Dr. X. Shang for writing the MATLAB

routine used to extract the resonant frequency and quality factor.

## REFERENCES

- [1] C. Gao, B. Hu, I. Takeuchi, K. S. Chang, X.-D. Xiang, and G. Wang, "Quantitative scanning evanescent microwave microscopy and its applications in characterization of functional materials libraries," *Meas. Sci. Technol.*, vol. 16, no. 1, p. 248, Dec. 2004.
- [2] Z. Wei *et al.*, "Quantitative theory for probe-sample interaction with inhomogeneous perturbation in near-field scanning microwave microscopy," *IEEE Trans. Microw. Theory Techn.*, vol. 64, no. 5, pp. 1402–1408, May 2016.
- [3] T. Monti, P. Iezzi, M. Farina, and S. W. Kingman, "Full electromagnetic simulation of a scanning microwave microscope for quantitative estimation of material properties," in *Proc. 1st URSI Atlantic Radio Sci. Conf. (URSI AT-RASC)*, Gran Canaria, Spain, May 2015, p. 1.
- [4] T. Monti, A. Di Donato, D. Mencarelli, G. Venanzoni, A. Morini, and M. Farina, "Near-field microwave investigation of electrical properties of graphene-ITO electrodes for LED applications," *J. Disp. Technol.*, vol. 9, no. 6, pp. 504–510, Jun. 2013.
- [5] S. Safari, A. Castellazzi, and P. Wheeler, "Experimental and analytical performance evaluation of SiC power devices in the matrix converter," *IEEE Trans. Power Electron.*, vol. 29, no. 5, pp. 2584–2596, May 2014.
- [6] P. M. Meaney, A. P. Gregory, N. R. Epstein, and K. D. Paulsen, "Microwave open-ended coaxial dielectric probe: Interpretation of the sensing volume re-visited," *BMC Med. Phys.*, vol. 14, no. 1, p. 3, Jun. 2014.
- [7] T. Monti, A. Tselev, O. Udoudo, I. N. Ivanov, C. Dodds, and S. W. Kingman, "High-resolution dielectric characterization of minerals: A step towards understanding the basic interactions between microwaves and rocks," *Int. J. Mineral Process.*, vol. 151, pp. 8–21, Jun. 2016.
- [8] R. Meisels, M. Toifl, P. Hartlieb, F. Kuchar, and T. Antretter, "Microwave propagation and absorption and its thermo-mechanical consequences in heterogeneous rocks," *Int. J. Mineral Process.*, vol. 135, pp. 40–51, Feb. 2015.
- [9] M. F. Cordoba-Erazo and T. M. Weller, "Noncontact electrical characterization of printed resistors using microwave microscopy," *IEEE Trans. Instrum. Meas.*, vol. 64, no. 2, pp. 509–515, Feb. 2015.
- [10] S. M. Anlage, V. V. Talanov, and A. R. Schwartz, "Principles of near-field microwave microscopy," in *Scanning Probe Microscopy*, S. Kalinin and A. Gruverman, Eds. New York, NY, USA: Springer, 2007, pp. 215–253.
- [11] M. Tabib-Azar, D. Akinwande, G. Ponchak, and S. R. LeClair, "Novel physical sensors using evanescent microwave probes," *Rev. Sci. Instrum.*, vol. 70, no. 8, pp. 3381–3386, Aug. 1999.
- [12] J. D. Chisum and Z. Popovic, "Performance limitations and measurement analysis of a near-field microwave microscope for nondestructive and subsurface detection," *IEEE Trans. Microw. Theory Techn.*, vol. 60, no. 8, pp. 2605–2615, Aug. 2012.
- [13] M. Tabib-Azar and B. Sutapun, "Novel hydrogen sensors using evanescent microwave probes," *Rev. Sci. Instrum.*, vol. 70, no. 9, pp. 3707–3713, Sep. 1999.
- [14] C. Gao and X.-D. Xiang, "Quantitative microwave near-field microscopy of dielectric properties," *Rev. Sci. Instrum.*, vol. 69, pp. 3846–3851, Nov. 1998.
- [15] T. Zhang and M. Tabib-Azar, "Calculation and accurate measurement of capacitance of electrically small axi-symmetric microstructures near a probe tip," in *61st ARFTG Conf. Dig.*, Philadelphia, PA, USA, Jun. 2003, pp. 147–156.
- [16] R. Inoue, Y. Odate, E. Tanabe, H. Kitano, and A. Maeda, "Data analysis of the extraction of dielectric properties from insulating substrates utilizing the evanescent perturbation method," *IEEE Trans. Microw. Theory Techn.*, vol. 54, no. 2, pp. 522–532, Feb. 2006.
- [17] R. J. Barlow, *Statistics: A Guide to the Use of Statistical Methods in the Physical Sciences*. New York, NY, USA: Wiley, 1989, p. 100.
- [18] A. Shivola, *Electromagnetic Mixing Formulas and Applications*. London, U.K.: IET digital library, 1999.
- [19] W. R. Tinga, "Mixture laws and microwave-material interactions," *Prog. Electromagn. Eng.*, vol. 6, no. 1, pp. 1–40, 1992.
- [20] T. Wei, X.-D. Xiang, W. G. Wallace-Freedman, and P. G. Schultz, "Scanning tip microwave near-field microscope," *Appl. Phys. Lett.*, vol. 68, no. 24, pp. 3506–3508, Jun. 1996.
- [21] D. J. Barker, T. J. Jackson, P. M. Suherman, M. S. Gashinova, and M. J. Lancaster, "Uncertainties in the permittivity of thin films extracted from measurements with near field microwave microscopy calibrated by an image charge model," *Meas. Sci. Technol.*, vol. 15, no. 10, Aug. 2014, Art. no. 105601.
- [22] M. J. Lancaster, *Passive Microwave Device Applications of High-Temperature Superconductors*. Cambridge, U.K.: Cambridge Univ., 1997, ch. 4.
- [23] P. Petersan and S. M. Anlage, "Measurement of resonant frequency and quality factor of microwave resonators: Comparison of methods," *J. Appl. Phys.*, vol. 84, no. 6, pp. 3392–3402, Sep. 1998.
- [24] M. Tabib-Azar, D.-P. Su, A. Pohar, S. R. LeClair, and G. Ponchak, "0.4  $\mu\text{m}$  spatial resolution with 1 GHz ( $\lambda=30$  cm) evanescent microwave probe," *Rev. Sci. Instrum.*, vol. 70, no. 3, pp. 1725–1729, Mar. 1999.
- [25] D.-Y. Sheu, "Micro-spherical probes machining by EDM," *J. Micromech. Microeng.*, vol. 15, no. 1, pp. 185–189, Oct. 2004.
- [26] D. J. Barker, "Evaluation of microwave microscopy for dielectric characterisation," Ph.D. dissertation, School Electron., Elect. Comput. Eng., Univ. Birmingham, Birmingham, AL, USA, 2010.
- [27] S. Gevorgian, *Ferroelectrics in Microwave Devices, Circuits and Systems*. London, U.K.: Springer-Verlag, 2009, ch. 4, p. 118.
- [28] R. M. Tennant, *Science Data Book*. Edinburgh, U.K.: Oliver & Boyd, 1987.
- [29] G. Q. Jiang, W. H. Wong, E. Y. Raskovich, W. G. Clark, W. A. Hines, and J. Sanny, "Measurement of the microwave dielectric constant for low-loss samples with finite thickness using open-ended coaxial-line probes," *Rev. Sci. Instrum.*, vol. 64, no. 6, pp. 1622–1626, Jun. 1993.
- [30] P. J. Goodhew and F. J. Humphreys, *Electron Microscopy and Analysis*. London, U.K.: Taylor & Francis, 1988, ch. 5.
- [31] L. Tsang and J. A. Kong, "Scattering of electromagnetic waves from random media with strong permittivity fluctuations," *Radio Sci.*, vol. 16, no. 3, pp. 303–320, May/Jun. 1981.

**Tamara Monti**, photograph and biography not available at the time of publication.

**Ofonime B. Udoudo**, photograph and biography not available at the time of publication.

**Kevin A. Sperin**, photograph and biography not available at the time of publication.

**Chris Dodds**, photograph and biography not available at the time of publication.

**Sam W. Kingman**, photograph and biography not available at the time of publication.

**Timothy J. Jackson**, photograph and biography not available at the time of publication.